

NAR[®] LITTER



MAINTENANCE INSPECTION

Although all NAR[®] litters are designed for use in rugged conditions, they still may require regular preventative maintenance.

INTENDED USE

The litter beds are intended for casualty extraction.

PRODUCT DESCRIPTION

The litter beds are designed to provide a platform and handles for rapid casualty support and extraction.

INDICATIONS FOR USE

The litter beds are indicated for use to provide a patient platform for casualty extraction in field environments.

CONTRAINDICATIONS

None

USER GROUPS

The litter beds are intended to be used by emergency responders and medical professionals including military medics and Emergency Medical Services (EMS) professionals.

PRODUCTS WITH A **MISSION[®]**

ZZ-0234 • REV030526



NORTH AMERICAN RESCUE[®]

www.NARescue.com • 888.689.6277

35 Tedwall Court • Greer, SC 29650 • USA

NAR[®] LITTER



SAFETY INSPECTION

PRODUCTS WITH A MISSION[®]

To maintain service life and high performance, all belts, hinges, handles, and other materials must be inspected before use, and replaced if any defective conditions are found.

- Examine each belt. Beginning at one end, holding the body side of the belt toward you, grasp the belt with your hands 6 to 8 inches apart.
- Bend the belt into an inverted U shape. This procedure will allow you to locate damaged fibers or cuts more readily. Watch for frayed edges, broken fibers, pulled stitches, cuts, or chemical damage.
- Special attention should be given to the attachment of buckles where attached to webbing. Note any unusual wear, frayed or cut fibers, or distortion of buckles.
- Look for frayed or broken fiber strands that appear as tufts of fiber on the webbing surface.
- Inspect frame, bars, and welds points for cracks, distortions, or stress marks.
- Check pins to ensure they are in place and have fasteners attached properly.
- Inspect handles for cracks, distortions, or stress marks.
- Examine litter bed for general rips, tears, and thread damage. Also look for signs of damage caused by heat, chemicals, or molten metal.
- If deemed necessary lubrication can be applied to all joints and moving parts. It is suggested that a molybdenum disulfide grease or equivalent be applied from an aerosol can. Any excess lubrication should be wiped clean of all external surfaces before placing litter back into service.



NORTH AMERICAN RESCUE[®]

www.NARescue.com • 888.689.6277