


KING LTS-D™ MILITARY Quick Reference Insert

*Refer to full IFU for detailed information and user tips.

 **NORTH AMERICAN RESCUE™**
INCORPORATED
35 Tedwall Ct. • Greer, SC 29650
1.888.689.6277 www.NARescue.com

3XSERIES™
COMPOSITE

1. Apply chin lift, if not contraindicated by C-spine precautions or patient position, and introduce KING LTS-D into corner of mouth, with blue orientation line facing laterally.



2. Advance tip under base of tongue, while rotating tube back to midline.



3. Without exerting excessive force, advance tube until base of connector is aligned with teeth or gums.



4. Inflate cuffs using the maximum volume of the syringe provided.



5. Attach resuscitator bag. While gently bagging, simultaneously withdraw the airway until ventilation is easy and free flowing (large tidal volume with minimal airway pressure). If remote ventilation is needed, use the flexible extension tubing to connect the KING LTS-D to the resuscitator bag.




- or -



6. If necessary, add additional volume to cuff to maximize seal of the airway.


7. For gastric access, note the length of gastric tube to be inserted using the black marks as references. Remove gastric diverter then lubricate the gastric tube prior to inserting it to the noted depth. Confirm placement in stomach. Use the least amount of suction that effectively decompresses the stomach.



- OVER -


KING LTS-D™ MILITARY Quick Reference Insert

*Refer to full IFU for detailed information and user tips.


 **NORTH AMERICAN RESCUE™**
INCORPORATED
35 Tedwall Ct. • Greer, SC 29650
1.888.689.6277 www.NARescue.com

3XSERIES™
COMPOSITE

1. Apply chin lift, if not contraindicated by C-spine precautions or patient position, and introduce KING LTS-D into corner of mouth, with blue orientation line facing laterally.



2. Advance tip under base of tongue, while rotating tube back to midline.



3. Without exerting excessive force, advance tube until base of connector is aligned with teeth or gums.



4. Inflate cuffs using the maximum volume of the syringe provided.



5. Attach resuscitator bag. While gently bagging, simultaneously withdraw the airway until ventilation is easy and free flowing (large tidal volume with minimal airway pressure). If remote ventilation is needed, use the flexible extension tubing to connect the KING LTS-D to the resuscitator bag.




- or -



6. If necessary, add additional volume to cuff to maximize seal of the airway.

7. For gastric access, note the length of gastric tube to be inserted using the black marks as references. Remove gastric diverter then lubricate the gastric tube prior to inserting it to the noted depth. Confirm placement in stomach. Use the least amount of suction that effectively decompresses the stomach.



- OVER -

KING LTS-D™ MILITARY Quick Reference Insert

*Refer to full IFU for detailed information and user tips.


 **NORTH AMERICAN RESCUE™**
INCORPORATED
35 Tedwall Ct. • Greer, SC 29650
1.888.689.6277 www.NARescue.com

3XSERIES™
COMPOSITE

1. Apply chin lift, if not contraindicated by C-spine precautions or patient position, and introduce KING LTS-D into corner of mouth, with blue orientation line facing laterally.



2. Advance tip under base of tongue, while rotating tube back to midline.



3. Without exerting excessive force, advance tube until base of connector is aligned with teeth or gums.



4. Inflate cuffs using the maximum volume of the syringe provided.



5. Attach resuscitator bag. While gently bagging, simultaneously withdraw the airway until ventilation is easy and free flowing (large tidal volume with minimal airway pressure). If remote ventilation is needed, use the flexible extension tubing to connect the KING LTS-D to the resuscitator bag.




- or -



6. If necessary, add additional volume to cuff to maximize seal of the airway.


7. For gastric access, note the length of gastric tube to be inserted using the black marks as references. Remove gastric diverter then lubricate the gastric tube prior to inserting it to the noted depth. Confirm placement in stomach. Use the least amount of suction that effectively decompresses the stomach.



- OVER -


KING LTS-D™ MILITARY Quick Reference Insert

*Refer to full IFU for detailed information and user tips.



ADDITIONAL PRODUCT KIT COMPONENTS

- Gastric Tube
- Extension
- 60cc Syringe
- Sterile Lubricant



SIZING INFORMATION

Size	Product	Criteria	Color	Inflation Volumes
3	KLTS-D	4-5 ft (122-155 cm)	Yellow	40-55 ml
4	KLTS-D	5-6 ft (155-180 cm)	Red	50-70 ml
5	KLTS-D	greater than 6 ft (>180 cm)	Purple	60-80 ml

KING LTS-D: For size 3, 4, and 5 the OD is 18 mm and the ID of the Ventilatory Lumen is equivalent to 10 mm. Gastric Tube Size <18 Fr


830-106 07/09

- OVER -

864-675-9800


KING LTS-D™ MILITARY Quick Reference Insert

*Refer to full IFU for detailed information and user tips.



ADDITIONAL PRODUCT KIT COMPONENTS

- Gastric Tube
- Extension
- 60cc Syringe
- Sterile Lubricant



SIZING INFORMATION

Size	Product	Criteria	Color	Inflation Volumes
3	KLTS-D	4-5 ft (122-155 cm)	Yellow	40-55 ml
4	KLTS-D	5-6 ft (155-180 cm)	Red	50-70 ml
5	KLTS-D	greater than 6 ft (>180 cm)	Purple	60-80 ml

KING LTS-D: For size 3, 4, and 5 the OD is 18 mm and the ID of the Ventilatory Lumen is equivalent to 10 mm. Gastric Tube Size <18 Fr


830-106 07/09

- OVER -

864-675-9800


KING LTS-D™ MILITARY Quick Reference Insert

*Refer to full IFU for detailed information and user tips.



ADDITIONAL PRODUCT KIT COMPONENTS

- Gastric Tube
- Extension
- 60cc Syringe
- Sterile Lubricant



SIZING INFORMATION

Size	Product	Criteria	Color	Inflation Volumes
3	KLTS-D	4-5 ft (122-155 cm)	Yellow	40-55 ml
4	KLTS-D	5-6 ft (155-180 cm)	Red	50-70 ml
5	KLTS-D	greater than 6 ft (>180 cm)	Purple	60-80 ml

KING LTS-D: For size 3, 4, and 5 the OD is 18 mm and the ID of the Ventilatory Lumen is equivalent to 10 mm. Gastric Tube Size <18 Fr

830-106 07/09

- OVER -

864-675-9800